



**SOP: Biotinidase (BTD)  
Enzyme Activity Assay Kit  
Cat. No. CS14**

**ID: SOPPCS14  
Version: V2.0  
Date: 05Feb26  
Pages: 13**

## Standard Operating Procedure:

### Biotinidase (BTD) Enzyme Activity Assay Kit Cat. No. CS14

**SOP ID:** SOPPCS14  
**Version Number:** V2.0  
**Version Date:** 05 February 2026  
**Product Name:** Biotinidase (BTD) Enzyme Activity Assay Kit  
**Matrices:** Dried Blood Spot

<b>Authored By:</b> Olive Lee white Bioanalytical Technician Enfanos, LLC.	 <i>Olive Lee White</i> F8940F4C7FA34049A98114DD1A9A89B2	Signed by Olive Lee White Date: 06-Feb-26 I am the author of this document 06-Feb-26   11:48:30 AM PST
<b>Reviewed By:</b> Zackary M. Herbst Chief Science Officer Enfanos, LLC.	 <i>Zackary M. Herbst</i> 95BFF6AAC0DB48458BA48EAB3EB25F25	Signed by Zackary M. Herbst Date: 06-Feb-26 I have reviewed this document 06-Feb-26   12:36:01 PM PST
<b>Approved By:</b> Hamid Khaledi Chief Executive Officer Enfanos, LLC.	 <i>Hamid Khaledi</i> 61319B2C6B7745E88E51504BBC9F42F9	Signed by Hamid Khaledi Date: 06-Feb-26 I approve this document 06-Feb-26   4:00:34 PM PST



## Table of Contents

<b>Section 1. Materials and Reagents.....</b>	<b>3</b>
<b>Table 1.1: Reagents and Standards.....</b>	<b>3</b>
<b>Table 1.2: Equipment and Consumables .....</b>	<b>4</b>
<b>Section 2. Preparation of the Assay Cocktail.....</b>	<b>5</b>
<b>Figure 2.1 Components of the BTD Enzyme Activity Assay Kit.....</b>	<b>5</b>
<b>2.1 Instructions for Preparation of Assay Cocktail.....</b>	<b>5</b>
<b>2.2 Storage and Stability of Reagents.....</b>	<b>6</b>
<b>Section 3. Recipes for Other Stocks and Buffers .....</b>	<b>7</b>
<b>3.1 Quench Solvent (1:1 (v:v) Methanol:Ethyl Acetate) .....</b>	<b>7</b>
<b>3.2 NaCl Solution (0.5 M Sodium Chloride in Water) .....</b>	<b>7</b>
<b>3.3 Reconstitution Solvent (1:1 (v:v) Water:Acetonitrile).....</b>	<b>7</b>
<b>Section 4. Step-by-Step Method.....</b>	<b>8</b>
<b>Figure 4.1 Enzyme Assay Scheme.....</b>	<b>8</b>
<b>Section 5. Recommended Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS) Method.....</b>	<b>10</b>
<b>Table 5.1: Materials and Reagents.....</b>	<b>10</b>
<b>5.1 LC Method.....</b>	<b>11</b>
<b>Table 5.2: Gradient Program.....</b>	<b>11</b>
<b>5.2 MRM Method .....</b>	<b>12</b>
<b>Table 5.3: MRM Method .....</b>	<b>12</b>
<b>5.3 Formula for Specific Activity .....</b>	<b>12</b>
<b>Section 6. References.....</b>	<b>13</b>
<b>Section 7. Document Audit Trail.....</b>	<b>13</b>



## Section 1. Materials and Reagents

**Table 1.1: Reagents and Standards**

<b>Material</b>	<b>Vendor</b>	<b>Vendor Reference</b>
Biotinidase (BTD) Enzyme Activity Assay Kit	Enfanos	CS14
MilliQ Purified Water (18.2 MΩ-cm resistivity), or Equivalent	Various	n/a
LC-MS Grade Ethyl Acetate* (CAS: 141-78-6)	JT Baker	9828-03
HPLC Grade Methanol (CAS: 67-56-1)	Fisher	A452
HPLC Grade Acetonitrile (CAS: 75-05-8)	Fisher	A998
Sodium Chloride (ACS Reagent Grade) (CAS: 7647-14-5)	Sigma Aldrich	S9888
N,N-Dimethylformamide (DMF) (CAS: 68-12-2)	Sigma Aldrich	319937 (or equivalent)

*\*Note: Trace amounts of oxidizers (e.g., peracetic acid) are present in some grades of Ethyl Acetate due to the manufacturing process. We suggest using the LC-MS grade of Ethyl Acetate available from J.T. Baker (cat. no. 9828-03), which has been found to be free of this issue. At this point, we cannot guarantee that Ethyl Acetate from this source will always be free of this issue. If the Ethyl Acetate quality is poor, the signal intensity and response factor could be impacted.*



**SOP: Biotinidase (BTD)  
Enzyme Activity Assay Kit  
Cat. No. CS14**

**ID: SOPPCS14  
Version: V2.0  
Date: 05Feb26  
Pages: 13**

**Table 1.2: Equipment and Consumables**

<b>Material</b>	<b>Vendor</b>	<b>Vendor Reference</b>
MilliQ Water Purification System	MilliQ	Any System
3.2 mm (1/8") paper punch	Any	Any
Parafilm	Any	Any
Deep-well 96-well plate (polypropylene)	Costar	3959
Silicone mat for 96-well plate	Costar	3080
(optional) 1.5 mL tube with snap cap (polypropylene)	VWR	89000-028
96-well microplate (polypropylene)*	Greiner Bio-One	651201
Zone-Free Sealing Film for 96-well microplate	Sigma	Z721646
Incubator with orbital shaker	Any	Any
Pipette, Single Channel, 2-20 $\mu$ L (or equivalent)	Any	Any
Pipette, Single Channel, 100-1000 $\mu$ L (or equivalent)	Any	Any
Pipette, Multichannel, 20-200 $\mu$ L (or equivalent)	Any	Any
P10, P200, and P1000 Pipette tips	Any	Any
Calibrated Hamilton Syringes for accurate small volume measurement	Hamilton	Any
Centrifuge for 96-well plate (swinging bucket rotor)	Any	Any
96-Well plate nitrogen jet manifold (for use at ambient temperature)	Any	Any
2-8°C Refrigerator (Explosion Proof Recommended)	Any	Any
$\leq$ -10°C Freezer	Any	Any
Fume Hood	Any	Any
Mass Spectrometer (see Table 5.1)	Any	Any

\* Note: Choose autosampler vials or 96-well autosampler plates that are compatible with the UPLC autosampler that will be used for analysis.

## Section 2. Preparation of the Assay Cocktail




Figure 2.1 Components of the BTD Enzyme Activity Assay Kit

### 2.1 Instructions for Preparation of Assay Cocktail

The S/IS Vial(s) (A) and Assay Buffer (B) are provided by Enfanos as a part of the Biotinidase (BTD) Enzyme Activity Assay Kit (Cat. No. CS14). Enzymatic Product (C) is provided as needed.

**Step 1.** Allow the Assay Buffer (B) and S/IS vial (A) to warm to room temperature.

**Step 2.** Next, add 15  $\mu\text{L}$  of DMF to the S/IS Mixture vial (for 50 assays) (A) using a calibrated Pipette and vortex until all substances are dissolved by eye. Then, add 1.485 mL of the Assay Buffer (B) to the S/IS Mixture vial (for 50 assays) (A) using a calibrated Pipette and vortex until all substances are dissolved. *For other vial sizes, reconstitution volume with DMF and Assay Buffer is printed on the S/IS Mixture Vial.*

Vortex the vial and allow to sit at room temperature for at least 30 minutes. After 30 minutes, vortex vial again and visually inspect with light behind it: a nearly clear solution should be obtained (a small amount of cloudiness may be noticed if the solution is held up to a light source).

The resulting Assay Cocktail contains 200  $\mu\text{M}$  BTD-S and 1  $\mu\text{M}$  BTD-IS. If you have question about the composition of the resulting Assay Cocktail, please contact us at [contact@enfanos.com](mailto:contact@enfanos.com). Additional components (detergent, inhibitors, activators, buffer salt, etc.) are non-hazardous and are not listed.



**SOP: Biotinidase (BTD)  
Enzyme Activity Assay Kit  
Cat. No. CS14**

**ID: SOPPCS14  
Version: V2.0  
Date: 05Feb26  
Pages: 13**

**Step 3.** To prepare a 200  $\mu$ M methanolic solution of the Enzymatic Product for LC-MS/MS tuning or standard curve preparation, add 5  $\mu$ L of methanol per 1 nmole of Enzymatic Product in the Enzymatic Product vial (**C**) using a Hamilton Syringe. Vortex to mix. Prepare additional dilutions as needed in 1:1 (v:v) Water:Acetonitrile (Reconstitution Solvent, see Section 3.3). The molecular weight of Enzymatic Product is 282.4 g/mol.

## 2.2 Storage and Stability of Reagents

Formal stability data for long-term storage of S/IS Vials (**A**) and Enzymatic Product Vials (**C**) are not available. We prefer to refrain from giving artificially short expiration dates, which would lead to wasteful discarding of reagents. However, our knowledge of chemical stability leads us to suggest that these solids are expected to be indefinitely stable when stored at  $\leq -10^{\circ}\text{C}$  in a jar containing desiccant (desiccants should contain a colored moisture indicator and should be replaced when spent).

The Assay Buffer (**B**) should be stored at 2-8 $^{\circ}\text{C}$ .

The Assay Cocktail (prepared in Section 2.1, Step 2) is stable frozen at  $\leq -10^{\circ}\text{C}$  for at least 1 month. Before storing the Assay Cocktail, wrap the cap tightly with ParaFilm. It can be frozen and thawed before use at least 2 times.



**SOP: Biotinidase (BTD)  
Enzyme Activity Assay Kit  
Cat. No. CS14**

**ID: SOPPCS14  
Version: V2.0  
Date: 05Feb26  
Pages: 13**

## Section 3. Recipes for Other Stocks and Buffers

### 3.1 Quench Solvent (1:1 (v:v) Methanol:Ethyl Acetate)

To prepare 500 mL of Quench Solvent, combine 250 mL of Methanol (HPLC Grade) and 250 mL of Ethyl Acetate (LC-MS Grade, see note in Table 1.1) in a borosilicate glass solvent bottle and mix well. Store at room temperature for up to 6 months.

### 3.2 NaCl Solution (0.5 M Sodium Chloride in Water)


To prepare 500 mL of NaCl Solution, weigh 14.61 g of Sodium Chloride (ACS Reagent Grade) and transfer to a borosilicate glass or polypropylene solvent bottle with 500 mL of MilliQ Purified Water (or equivalent). Mix well until all solids have dissolved. Store at 2-8°C for up to 12 months.

### 3.3 Reconstitution Solvent (1:1 (v:v) Water:Acetonitrile)

To prepare 500 mL of Reconstitution Solvent, combine 250 mL of HPLC Grade Acetonitrile and 250 mL of MilliQ purified water (or equivalent) in a borosilicate glass or polypropylene solvent bottle and mix well. Store at room temperature for up to 6 months.

## Section 4. Step-by-Step Method

Figure 4.1 Enzyme Assay Scheme



**Step 1.** Add one 3.2 mm DBS punch to a well in a 96-well deep-well plate (alternatively, use a 1.5 mL polypropylene microcentrifuge tube) and add 30  $\mu\text{L}$  of the **Assay Cocktail** (prepared in Section 2.1, Step 2).

Seal the plate with a silicone 96-well plate mat (or cap the tube). Spin the plate (or tube) in a centrifuge at 200g for 10-20 seconds to bring all the liquid to the bottom, then incubate on an orbital shaker set to shake appropriately\* for 3 hours at 37°C. Record the time when the plate is put on the incubator. This is the start time for the enzyme activity assay incubation.

\* Note: The exact shaking setting will vary depending on the equipment used. The laboratory should determine the optimal shaking speed for their incubator/shaker as part of method implementation.

**Step 2.** After incubation, spin the plate (or tube) in a centrifuge at 200g for 10-20 seconds to bring all the liquid to the bottom, then quench the reaction by addition of 100  $\mu\text{L}$  of Quench Solvent (see Section 3.1), 400  $\mu\text{L}$  of LC-MS Grade Ethyl Acetate (see note in Table 1.1), and 200  $\mu\text{L}$  of NaCl Solution (see Section 3.2). Pipet the mixture up and down 10 times to mix well.

Note: only polypropylene or glass materials should be used to dispense Ethyl Acetate, as the solvent dissolves most plastics. Work with Ethyl Acetate should be done in a fume hood.

Record the time when the Quench Solvent is added: this is the end time for the enzyme activity assay incubation.



**SOP: Biotinidase (BTD)  
Enzyme Activity Assay Kit  
Cat. No. CS14**

**ID: SOPPCS14  
Version: V2.0  
Date: 05Feb26  
Pages: 13**

Cover the plate with the silicone 96-well plate mat (or cap the tube) and spin in a centrifuge at 3000g for 5 minutes at ambient temperature to completely separate the aqueous (bottom) and Ethyl Acetate (top) layers.

**Step 3.** After centrifugation, transfer 200  $\mu$ L of the top Ethyl Acetate layer to a 96-well polypropylene microplate appropriate for the autosampler you will use.

Evaporate the solvent from each well at room temperature under a nitrogen jet 96-well manifold evaporator.

*\* Note: If the plate must be stored overnight prior to LC-MS/MS analysis, cover with zone-free sealing film and store at  $\leq -10^{\circ}\text{C}$ . On the day of LC-MS/MS analysis, allow the plate to warm to room temperature and proceed to Step 4.*

**Step 4.** On the same day as LC-MS/MS analysis, add 200  $\mu$ L of Reconstitution Solvent (see Section 3.3) to each well. Mix with a multichannel Pipette by pipetting up and down 10x.

Seal the plate using zone-free sealing film, or as appropriate for your autosampler, and submit to LC-MS/MS analysis (see Section 5 for recommended method conditions).



**SOP: Biotinidase (BTD)  
Enzyme Activity Assay Kit  
Cat. No. CS14**

**ID:** SOPPCS14  
**Version:** V2.0  
**Date:** 05Feb26  
**Pages:** 13

## Section 5. Recommended Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS) Method

**Table 5.1: Materials and Reagents**

Material	Vendor	Vendor Reference
Optima Water (LC-MS) (CAS: 7732-18-5)	Fisher	W6500
Optima Acetonitrile (LC-MS) (CAS: 75-05-8)	Fisher	A955
Optima Isopropanol (LC-MS) (CAS: 67-63-0)	Fisher	A461
Optima Formic Acid (LC-MS) (CAS: 64-18-6)	Fisher	A117
CSH C18 Column, 1.7 $\mu$ m, 2.1 X 50 mm*	Waters	186005296
CSH C18 Pre-column, 1.7 $\mu$ m, 2.1 X 5 mm*	Waters	186005303
Acquity 2D UPLC System*	Waters	Or Equivalent
Triple Quad Mass Spectrometer	Waters	Or Equivalent

\* Note: While Enfanos recommends use of Waters equipment for LC-MS/MS analysis, equivalent instrumentation may be used for analysis of our enzyme activity assay products and internal standards.



### 5.1 LC Method

- **Solvent A.** 0.1% Formic Acid in 70:30 (v:v) Water:Acetonitrile
- **Solvent B.** 0.1% Formic Acid in 65:35 (v:v) Isopropanol:Acetonitrile
- **Weak Needle Wash.** 0.1% Formic Acid in 90:10 (v:v) Water: Acetonitrile
- **Strong Needle Wash.** 0.1% Formic Acid in Acetonitrile
- **Column Temperature.** 55°C

**Table 5.2: Gradient Program**

Time (min)	Flow (mL/min)	Solvent B (%)
initial	0.8	0.5
0.75	0.8	22
1.00	0.8	28
1.50	0.8	40
1.80	0.8	100
2.15	0.8	100
2.20	0.8	0.5

\* Note: To minimize contamination of the ESI source, we advise to divert the LC flow to waste except in the retention time region (see Table 5.3) where the product and internal standard elute (~1.20 min).

\*\* Note: In-source breakdown of the BTD Substrate may result in an interfering peak in the BTD Product MRM trace chromatogram. To prevent this interference, special attention must be paid to ensure baseline separation between the substrate and the product/internal standard peaks. If baseline separation is not achieved on the first try, modify the LC gradient. The BTD-S MRM is given in Table 5.3 for this purpose. Reach out to us with any questions ([contact@enfanos.com](mailto:contact@enfanos.com)).



**SOP: Biotinidase (BTD)  
Enzyme Activity Assay Kit  
Cat. No. CS14**

**ID: SOPPCS14  
Version: V2.0  
Date: 05Feb26  
Pages: 13**

## 5.2 MRM Method

MRM Transitions (see Table 5.3) are measured in ESI positive mode.

*\* Note: Method developed for analysis on Waters XEVO TQD Instrument (or instrument with equivalent sensitivity). If a higher sensitivity instrument, such as a Waters XEVO TQ-S Micro is being used we recommend adjusting sample preparation so the injected solution contains less moles of analyte or detuning the collision energy of your analytes so the internal standard peak areas are approximately 10,000-100,000 AUC. This will ensure linearity throughout the analytical range*

**Table 5.3: MRM Method**

Analyte	Parent (m/z)	Product (m/z)	Retention Time (min)
BTD Product	283.3	91.0	1.20
BTD Internal Standard	287.2	91.0	1.20
BTD Substrate	509.2	120.1	1.30

*\* Note: Transitions, cone voltage, collision energy, and retention times need to be confirmed or determined by tuning on your LC-MS/MS instrument in the usual way.*

## 5.3 Formula for Specific Activity

$$\text{BTD Specific Activity } (\mu\text{mol/h/L}) = \frac{P}{IS} \times \frac{0.00003 \mu\text{mole}}{3 \text{ h} \times 3.2 \times 10^{-6} \text{ L}}$$

To obtain the enzymatic activity, multiply the product-to-internal standard peak area ratio by the micromoles of internal standard in the assay (0.00003  $\mu\text{mol}$ ). This number is then divided by the incubation time (3 hours) the liters of blood in a 3.2 mm DBS punch ( $3.2 \times 10^{-6}$  L).

*\* Note: To calculate the incubation time, subtract the enzyme activity assay end time (see Section 4, Step 2) from the enzyme activity assay start time (see Section 4, Step 1).*

*\*\* Note: This procedure describes the preparation and handling of research-use-only materials. It does not establish or support clinical testing, diagnostic use, or patient result interpretation. Any use of this product in laboratory developed tests or clinical applications is the sole responsibility of the user. Enfanos does not validate or support clinical or diagnostic use.*

*This Enfanos product is intended for research use only (RUO). It is not intended for clinical or diagnostic use.*



**SOP: Biotinidase (BTD)**  
**Enzyme Activity Assay Kit**  
**Cat. No. CS14**

**ID:** SOPPCS14  
**Version:** V2.0  
**Date:** 05Feb26  
**Pages:** 13

## Section 6. References

Hong, X., Kumar, A. B., Scott, C. R., & Gelb, M. H. (2018). Multiplex tandem mass spectrometry assay for newborn screening of X-linked adrenoleukodystrophy, biotinidase deficiency, and galactosemia with flexibility to assay other enzyme assays and biomarkers. *Molecular genetics and metabolism*, 124(2), 101-108.

Hong, X., Sadilek, M., & Gelb, M. H. (2020). A highly multiplexed biochemical assay for analytes in dried blood spots: application to newborn screening and diagnosis of lysosomal storage disorders and other inborn errors of metabolism. *Genetics in Medicine*, 22(7), 1262-1268.

## Section 7. Document Audit Trail

*Previous versions of this SOP are stored internally by Enfanos, LLC. If you require a copy of a previous version for comparison to this most recent version, please contact us at [contact@enfanos.com](mailto:contact@enfanos.com)*